

DSH  
JPD

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 6/12/2015

Instrument Serial Number: 010626	As of 12-Jun-15
----------------------------------	-----------------

Start Date	End Date	Agency Name
11-Jun-15		DFS Central Lab
15-Jan-15	11-Jun-15	Nottoway County SO
16-Dec-14	15-Jan-15	DFS Central Lab
01-Oct-14	16-Dec-14	Rappahannock Shenandoah Warren Regional Jail
01-Oct-14	01-Oct-14	DFS Central Lab
15-Jul-14	01-Oct-14	Warren County SO
17-Jun-14	15-Jul-14	DFS Central Lab
03-Dec-08	17-Jun-14	Patrick County Sheriff's Office
04-Aug-08	03-Dec-08	DFS Central Lab

DBH  
JPD

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010626 Worksheet Start Date 6/11/2015  
Location Nottoway CO SO  
Address 266 W Courthouse Rd. Nottoway VA 23995  
DFS Technician Donald Hall License No. 21250

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 749 Reference Barometer (mm HG) 749  
Reference Barometer(RB)Serial # 009109 RB Calibration Due 7/28/2015

## Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.296	0.287	0.304	0.296	0.295	0.295
Precision		sample min	sample max			
0.001		0.295	0.296			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.099	0.096	0.102	0.097	0.098	0.097
Precision		sample min	sample max			
0.001		0.097	0.098			

Dry gas standard Lot No. (with tank no.)


AG407801-12

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:

Instrument Serial Number

010626

Certification Date

06/12/2015

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Jondal B Hall

Date

6/12/15

Issuing Analyst

Jennie Dufly  
JMB

Date

6/12/15

DBH  
JPD




COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>HALL, DONALD, B</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>21250</b>	LICENSE EXPIRES <b>10/01/2016</b>	DATE TEST CONDUCTED <b>06/11/2015</b>
TEST EQUIPMENT NUMBER <b>010626</b>		

RESULTS: TIME SAMPLE TAKEN 10:40 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20 \_\_\_\_\_.

BREATH TEST OPERATOR
----------------------

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_ OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 11 Jun 2015 at 11:07

DBH  
JPD

## Test Results

Instrument Serial Number 010626

Test # 000553 Subject Test

Test Location 1 Department of

Test Date 11 Jun 2015

Test Location 2 Forensic Science

Test Time 10:34

Remote/Local Local

Test Location 3

System Check Passed

Operator's Last Name HALL

Agency DFS Central Lab

Card Serial Number 121250

Subject's Last Name INSTRUMENT

Subject's Middle Initial

Driver's License Number

Driver's License State

Operator's First Name DONALD

Effective Date 10/01/2014

Subject's Date of Birth 00/00/0000

Driver's License Expiration 00/00/0000

Court Name DFS

Operator's Middle Initial B

License Number 21250

Expiration Date 10/01/2016

Subject's Sex Male

End Date 11 Jun 2015

End Time 10:41

Result Time 10:40

Result Date 11 Jun 2015

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:34

Data Type BLK

Sample Value 0.000

Sample Time 10:35

Data Type CHK

Sample Value 0.098

Sample Time 10:35

Data Type BLK

Sample Value 0.000

Sample Time 10:37

Data Type SUBJ

Sample Value 0.000

Sample Time 10:38

Data Type BLK

Sample Value 0.000

Sample Time 10:39

Data Type SUBJ

Sample Value 0.000

Sample Time 10:40

Data Type BLK

Sample Value 0.000

Sample Time 10:41

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG407801-12

Standard Expiration Date 03/19/2016

Tank Pressure 948

Barometric Pressure 749 mmHg

Blow Sample Number 1 Blow Duration 3.53 sec

Blow Volume 1855 cc End-of-Blow Time 10:38

Blow Sample Number 2 Blow Duration 3.74 sec

Blow Volume 1947 cc End-of-Blow Time 10:40

Tamper Evident Stamp ff48eca0

Test Status Code 0

Test Status Success

**Intox EC/IR-II: Accuracy Check**

SBH  
JPD

*Department of Forensic Science*

Serial Number: 010626      Test Number: 554

Test Date: 06/11/2015      Test Time: 10:43 EDT

Dry Gas Target: 0.099

Lot Number: AG407801-12      Exp Date: 03/19/2016

Tank Pressure: 943 psi      Barometric Pressure: 749 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:43
CHK	0.097	10:44
BLK	0.000	10:46
CHK	0.098	10:46
BLK	0.000	10:48
CHK	0.097	10:49

Test Status: *Success*

Calibration CRC: A6B9F496

**Intox EC/IR-II: Accuracy Check**

DBH  
JPD

*Department of Forensic Science*

Serial Number: 010626      Test Number: 555

Test Date: 06/11/2015      Test Time: 10:51 EDT

Dry Gas Target: 0.296

Lot Number: AG506501-12      Exp Date: 03/06/2017

Tank Pressure: 325 psi      Barometric Pressure: 749 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:52
CHK	0.296	10:53
BLK	0.000	10:54
CHK	0.295	10:55
BLK	0.000	10:57
CHK	0.295	10:58

Test Status: *Success*

Calibration CRC: A6B9F496